

Certificate



Certificate no.

TA 50373180 01

License Holder:

TDK-Lambda Americas Inc.
401 Mile of Cars Way, Suite 325,
National City, CA 91950
USA

Manufacturing Plant:

ZTW1-LinPr-10049856 001

Test report no.: ZTW1-SCL 10049856 001

Client Reference: RD06-13022017-Co.TDK

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

| | | |
|---------------------------|---------------------------------|---|
| Type Designation | : CSS65A-XZZZZZ (TDK-Lambda) | 5 |
| X stands for | : 5, 12, 15, 19, 24, 28, 48, 54 | 1 |
| Z stands for | : a-z, A-Z, 0-9 | 1 |
| Rated Voltage | : AC 100-240V, 50-60Hz | |
| Rated Current | : 2-1A | |
| Output Voltage | : see Appendix | |
| Output Current | : see Appendix | |
| max. Ambient Temperature: | 50°C | |
| Protection Class | : I | |

Remark: 1. This risk management requirements of the standard were not addressed.
2. This certificate is valid for maximum 5 years.

7


Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

07/03/2017

Appendix to TÜV approved Certificate No.: TA 50373180

Certified Product : Switching Power Supply
 Report Number : 10049856 001
 Type Designation : CSS65A-XZZZZZ
 (X = 5, 12, 15, 19, 24, 28, 48, 54; Z = a-z, A-Z, 0-9)

| Model name | Output rating | | |
|----------------|---------------|------|----|
| | V | A | W |
| CSS65A-5ZZZZZ | +5 | 8 | 40 |
| CSS65A-12ZZZZZ | +12 | 5.42 | 65 |
| CSS65A-15ZZZZZ | +15 | 4.34 | 65 |
| CSS65A-19ZZZZZ | +19 | 3.43 | 65 |
| CSS65A-24ZZZZZ | +24 | 2.71 | 65 |
| CSS65A-28ZZZZZ | +28 | 2.33 | 65 |
| CSS65A-48ZZZZZ | +48 | 1.36 | 65 |
| CSS65A-54ZZZZZ | +54 | 1.21 | 65 |

Date: 7th March, 2017

Certification Body


 Dipl.-Ing. W. Hsu



Certificate



Certificate no.

TA 50373180 02

License Holder:

TDK-Lambda Americas Inc.
401 Mile of Cars Way, Suite 325,
National City, CA 91950
USA

Manufacturing Plant:

040-0001729524

Test report no.: ZTW1-RC 10049856 003

Client Reference: P01255431

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

Remark : This certificate is valid for maximum 5 years from last page.

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419045040

TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

21/11/2023

